

The 5014iQ and 5030iQ SHARP Particulate Monitors

Jeffrey Ambs
Sr. R&D Project Manager
Thermo Fisher Scientific



 The world leader in serving science

Introduction

Two models, both use beta attenuation for mass measurements

Model 5014iQ PM Monitor

- Beta attenuation monitor
- Provides reliable mass measurement
- Provides real-time beta measurement



Model 5030iQ SHARP Monitor


- Adds nephelometer for fast response




Operating Principal of SHARP Monitor

SHARP: Synchronized Hybrid Ambient Real-time Particulate


- Combines Beta Attenuation measurement with nephelometer
- Nephelometer is continuously corrected to mass using beta mass measurements




Model 5030iQ PM2.5 Comparison, Bakersfield, CA



PM10 Comparison: 5014 iQ Beta PM and 5030iQ SHARP



PM2.5 Comparison: 5014iQ Beta PM and 5030iQ SHARP




Appearance Redesign and Tape Position

Moved the filter tape access from the side to the front

Redesign to realize simpler operation and maintenance

Features:

1. Front door access for filter tape replacement and maintenance
2. Larger tape spool for up to 12-months operation
3. Window to view tape spool
4. Sample inlet X-Y position compatible with iSeries and legacy



Upgraded Heating Tube

Redesign the heating tube to achieve better heating performance.

Features:

1. Better heat preservation
2. Outer hard nylon shell to eliminate the risk of carbonization
3. Real-time temperature monitoring: temperature sensor tube outlet provides better control
4. Eliminates safety concern of overheating



Upgraded Beta Measurement Module

Photomultiplier Tube (PMT) beta detector

Upgraded to resolve issues with proportional counter:

1. Improved performance compared to previous detector
2. Elimination of high noise amplifier
3. Reduced start-up time: < 30 minutes compared to > 12 hours




- Test campaigns in China
 - Beijing
 - Shenzhen
- Test campaigns in US
 - Bakersfield
 - St. Louis
 - Ft. Collins
 - New Haven

Regulatory Approvals


- China – CCEP complete
- U.S. EPA Approvals complete
 - 5014iQ
 - PM-10 EQPM-1102-150
 - PM-2.5 EQPM-0609-183
 - 5030iQ
 - PM-2.5 EQPM-0609-184
 - PM-10 EQPM-0423-260

Beijing PM2.5 Test Results: Model 5030iQ SHARP



Beijing PM2.5 Test Results: Model 5030iQ SHARP





Beijing PM10 Test Results: Model 5030iQ Beta PM



Shenzhen PM2.5 Test Results: Model 5030iQ Beta PM



Bakersfield Test Results: Model 5030iQ SHARP





Precision	Data set mean, $\mu\text{g}/\text{m}^3$		Data set precision, $\mu\text{g}/\text{m}^3$		Relative precision (CV)	
	FRM	Candidate	FRM	Candidate	FRM	Candidate
Mean:	27.4	27.6	0.3	0.5	1.6%	2.6%
Maximum:	56.4	56.1	0.9	1.4	5.8%	10.3%
Minimum:	3.8	4.2	0.1	0.1	0.2%	0.2%
Candidate / FRM Ratio:	100.6%		149.0%		164.4%	
RMS Relative Precision for this site:					2.0%	3.8%
Test requirements - PM2.5 Class III:					10.0%	15.0%
Precision Test Results for site:					OK	PASS

Regression statistics		Slope ¹	Intercept ²	Correlation (r)
Statistics for this test site:		0.995	0.307	0.99639
Limits for PM2.5 Class III	Upper:	1.100	1.916	
	Lower:	0.900	-2.000	0.95000
Test Results (Pass/Fail):		PASS	PASS	PASS



East St. Louis Test Results: Model 5030iQ SHARP




Precision	Data set mean, $\mu\text{g}/\text{m}^3$		Data set precision, $\mu\text{g}/\text{m}^3$		Relative precision (CV)	
	FRM	Candidate	FRM	Candidate	FRM	Candidate
Mean:	8.1	8.4	0.3	0.5	3.4%	5.8%
Maximum:	14.3	14.6	0.9	1.2	9.6%	11.6%
Minimum:	4.0	3.8	0.1	0.1	0.5%	1.5%
Candidate / FRM Ratio:	103.4%		195.3%		171.1%	
RMS Relative Precision for this site:					3.9%	6.3%
Test requirements - PM2.5 Class III:					10.0%	15.0%
Precision Test Results for site:					OK	PASS

Regression statistics		Slope ¹	Intercept ²	Correlation (r)
Statistics for this test site:		0.933	0.814	0.97593
Limits for PM2.5 Class III	Upper:	1.100	2.000	
	Lower:	0.900	-1.117	0.93000
Test Results (Pass/Fail):		PASS	PASS	PASS




Ft. Collins Test Results: Model 5030iQ SHARP




Precision	Data set mean, µg/m ³		Data set precision, µg/m ³		Relative precision (CV)	
	FRM	Candidate	FRM	Candidate	FRM	Candidate
Mean:	7.7	7.8	0.4	0.6	5.4%	7.1%
Maximum:	20.1	20.9	0.8	3.6	9.4%	22.7%
Minimum:	3.8	3.3	0.0	0.0	0.2%	0.8%
Candidate / FRM Ratio:	101.2%		143.0%		131.6%	
RMS Relative Precision for this site:					5.9%	8.8%
Test requirements - PM2.5 Class III:					10.0%	15.0%
Precision Test Results for site:					OK	PASS

Regression statistics	Slope ¹	Intercept ²	Correlation (r)
Statistics for this test site:	1.033	-0.166	0.98334
Limits for PM2.5 Class III	Upper:	1.100	1.408
	Lower:	0.900	-2.000
Test Results (Pass/Fail):	PASS	PASS	PASS





New Haven Test Results: Model 5014iQ Beta PM




Precision	Data set mean, $\mu\text{g}/\text{m}^3$		Data set precision, $\mu\text{g}/\text{m}^3$		Relative precision (CV)	
	FRM	Candidate	FRM	Candidate	FRM	Candidate
Mean:	6.3	6.5	0.2	0.4	2.9%	7.7%
Maximum:	10.0	10.8	0.6	1.1	9.3%	30.0%
Minimum:	3.0	2.4	0.0	0.1	0.6%	0.7%
Candidate / FRM Ratio:	103.2%		248.4%		262.8%	
RMS Relative Precision for this site:					3.7%	10.2%
Test requirements - PM2.5 Class III:					10.0%	15.0%
Precision Test Results for site:					OK	PASS

Regression statistics		Slope ¹	Intercept ²	Correlation (r)
Statistics for this test site:		1.056	-0.154	0.97850
Limits for PM2.5 Class III	Upper:	1.100	1.112	
	Lower:	0.900	-2.000	0.93000
Test Results (Pass/Fail):		PASS	PASS	PASS



Instrument Comparison Results: Model 5030iQ SHARP



Thank You

